

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for Ki67

Obtained in run B22

19 Aug 2016

Immunostainer

Type: Gene Tech Genestainer

Primary antibody

Clone: GM001 Producer: Genemed

Product no. / lot no.: 61-0040 / 60525272

Diluent: Antibody diluent

Dilution factor: 1:120

Incubation time / temperature: 50 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker

Buffer: Std Tris-EDTA / EGTA pH 9

Heating time at max. temp.: 3 min. Maximum heating temp.: 25°C

Visualization system

Producer: Gene Tech

Product / no: GTVision / GK6005

Linker: None Incubation time polymer: 20 min. Incubation temperature: 25°C

Chromogen

Producer: Gene Tech

Product / no: DAB Detection Kit / GK6005

Incubation time / temperature: 5 min. / 25°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.