

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for OCT3/4

Obtained in run 53

06 Mar 2018

Immunostainer

Type: Gene Tech Genestainer

Primary antibody

Clone: C10 Producer: zeta

Product no. / lot no.: Z2104 / G0114

Diluent: Ab diluent

Dilution factor: 1:60

Incubation time / temperature: 50 min. / 25°C

Epitope retrieval, HIER

Device: Water bath

Buffer: Std Tris-EDTA / EGTA pH 9

Heating time at max. temp.: 15 min. Maximum heating temp.: 95°C

Visualization system

Producer: Gene Tech

Product / no: GTVision / GK6007

Linker: None Incubation time polymer: 20 min. Incubation temperature: 25°C

Chromogen

Producer: Gene Tech

Product / no: DAB Detection Kit / GK6007

Incubation time / temperature: 5 min. / 25°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.