

Recommended protocol for IgL

Obtained in run

21 Aug 2016

Immunostainer

Type: Dako Autostainer

Primary antibody

Clone: Polyclonal
Producer: Dako
Product no. / lot no.: A0193 / 081
Diluent: Antibody Diluent
Dilution factor: 1:4000
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Microwave oven
Buffer: Std Citrate pH 6
Heating time at max. temp.: 15 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Dako
Product / no: REAL EnVision Rabbit/Mouse / K5007 / K4065
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: DAB+ / K5007 / K4065
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.