

NORDIQC DATA FOR BREAST MARKERS

Antibody selection, protocols and controls

NordiQC Workshop, September 29th – October 1st 2021

Heidi Lykke Kristoffersen Staff Scientist NordiQC

AGENDA



- NordiQC results for selected markers
- Clones successful vs. less successful
- Tricky markers pitfalls

• iCAPS





	Purpose	Last run	Pass rate	No of lab
GATA3	<u>Breast</u> vs non-breast	Run 54, 2018	76% 🔶	245
Mammaglobin	<u>Breast</u> vs non-breast	Run 25, 2009	83%	23

Type I IHC tests

Type II IHC tests

GCDFP15 Breast vs non-breast Run 36, 2012 86% 131 263 CK5 Run 55, 2019 44% CIS vs invasive Run 50, 2017 78% 114 SMH <u>CIS</u> vs invasive Run 61, 2021 324 CIS vs invasive 79% p63 Ductal vs lobular E-Cadherin Run 53, 2018 89% 298 KI67 PI index Run B22, 2016 93% 409 ER Predictive for Tamoxifen Run B31, 2021 85% 378 PR Predictive for Tamoxifen Run B31, 2021 92% 377 HER2 IHC Predictive for Herceptin Run B31, 2021 362 92% Predictive for Herceptin HER2 BRISH Run H19, 2021 143 70% Predictive for Tecentriq 125 PD-L1 IC Run C9, 2021 69%

Scheduled for assessment within the next year



KEY-POINTS FOR BEST PROTOCOLS



- Clone selection
- RTUs "Plug and Play" or "Play and Plug"?
- Efficient HIER, preferable in an alkaline buffer
- Use of right detection system fit for purpose

Use of iCAPS

CLONE PERFORMANCE FOR SELECTED BREAST MARKERS



Marker	Successful clones (pass rate)	Less successful clones (pass rate)
GATA3	mAb L50-823 (77%), rmAb SP368 (100%)	mAb HG3-31 (0%)
СК5*	mAb XM26 (79%), rmAb SP27 (100%)	mAb D5/16 B4 (23%)
SMH	mAb SMMS1 (79%)	
p63	mAbs 4A4 (82%) & DAK-p63 (87%)	mAb 7JUL (10%)
E-Cadherin	mAbs NCH-38 (97%), 36 (94%) & 36B5 (95%)	rmAb EP700Y (25%)
KI67	mAb MIB-1 (90%), rmAb 30.9 (99%)	-
ER	rmAbs SP1 (89%) & EP1 (84%), mAb 6F11 (65%)	-
PR	mAbs 16 (90%) & PgR1294 (91%), rmAbs 1E2 (95%) & Y85 (100%)	-
HER2 IHC	rmAbs 4B5 (99%) & DG44 (100%), Dako pAb (85%)	mAb CB11 (50%)
PD-L1 IC	rmAb SP142 (89%)	Non-SP142
*for pitfalls: <u>see p</u>	pt for lung-markers	

ICAPS FOR SELECTED BREAST MARKERS



Marker	IHC critical assay performance controls Low expression	Negative tissue controls No expression		
GATA3	Tonsil: T-helper-cells in the T-zones and germinal centers.	Tonsil: B-cells, squamous epithelial cells, endothelial cells.	<u>Link</u>	
Mammaglobin	Skin: Epithelial cells of eccrine sweat glands.	Tonsil: All cell types.	<u>Link</u>	
GCDFP15	Skin: Epithelial cells of eccrine sweat glands.	Tonsil: All cell types.	<u>Link</u>	
CK5	Pancreas: Scattered epithelial cells of intercalated ducts.	Liver. All cell types.	<u>Link</u>	
Smooth MHCM	Tonsil: Follicular dendritic cells in germinal centers.	Tonsil: Epithelial cells.	<u>Link</u>	M Dog
p63	Placenta: Cytotrophoblastic cells.	Appendix: Epithelial- and smooth muscle cells.	<u>Link</u>	
E-Cadherin	Liver: Hepatocytes.	Appendix: Stromal cells, smooth muscle cells, endothelial cells.	<u>Link</u>	COLLE
KI67	Tonsil: B-cells in the light zones of the germinal centers.	Liver: Hepatocytes	<u>Lini</u> T	Door Co
ER	Tonsil: Squamous epithelial cells, T-cells in germinal centres.	Tonsil: B-cells in mantle zone and germinal centres.	Link	COOLCONTROL!!
PR	Cervix: Basal squamous epithelial cells.	Tonsil: All cells types (especially focus on lymphocytes in germinal centres).	Link	
PD-L1 IC	Tonsil: T-cells and macrophages in germinal centres.	Tonsil: Normal squamous epithelial cells, lymphocytes.	Link	



GATA3 – PITFALLS/POINTS OF ATTENTION

Table 1. Antibodies and assessment marks for GATA3, run 54								
Concentrated antibodies	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	Suff. OPS ²
mAb clone L50-823	38 77 4 2 1 1 1 1 1 1 2	Biocare Cell Marque BD Bioscience Zytomed Immunologic Zeta DBS DCS Lifespan Neomarkers Nordic Biosite Menapath	51	39	14	26	69%	80%
mAb clone HG3-31	2	Santa Cruz	0	0	0	2	-	-
rmAb clone EP368	3	Cell Marque	3	0	0	0	-	-
Unknown	2		1	0	0	1	-	-
Ready-To-Use antibodies								
mAb clone L50-823 760-4897	92	Ventana	65	19	6	2	91%	97%
mAb Clone L50-823 390M-17,18,10	9	Cell Marque	3	3	3	0	-	-
mAB Clone L50-823 PM 405AA	3	Biocare Medical	0	2	1	0	-	-
mAB clone L50-823 PM199	1	Path N Situ	0	0	0	1	-	-
mAB clone L50-823 MAB-0695	1	Maixin	0	1	0	0	-	-
mAB clone L50-823 MAD-000632QD	2	Master Diagnostica	0	0	1	1	-	-
Total	245		123	64	25	33	-	
Proportion			50%	26%	10%	14%	76%	

1) Proportion of sufficient stains (optimal or good).

Proportion of sufficient stains with optimal protocol settings only, see below.

Table 3. Proportion of optimal results for GATA3 for the most commonly used antibody as concentrate on the								
4 main IHC syst	ems*							
Concentrated	Dal	ko	Dal	(0	Ven	tana	Le	ica
antibodies	Autostainer Link /		ier Link / Omnis		BenchMark XT /		Bond III / Max	
	Classic				Ultra		1	
	TRS pH	TRS pH	TRS pH	TRS pH	CC1 pH	CC2 pH	ER2 pH	ER1 pH
	9.0	6.1	9.0	6.1	8.5	6.0	9.0	6.0
mAb clone	3/21**	-	9/16 (56%)	0/3	28/41	1/1	2/12	0/2
* Antibody concentr systems. ** (number of optin	ation applied as	listed above, l	HIER buffers and	l detection kits uffer)	used as prov	ided by the ve	endors of the r	espective

No RTU products for Dako and Leica users. Use of conc. format of mAb L50-823 can obtain optimal results.

Recommended protocol settings:

- HIER in an alkaline buffer

- 38% pass rate for 2-step detection systems (10% optimal)
- 83% pass rate for 3-step detection systems (56% optimal)

GATA3 – ICAPS





Tonsil

GATA3 – PITFALLS/POINTS OF ATTENTION





GATA-3 staining of TMA with 20 different neoplasias

3 of 3 breast carcinomas positive

1 of 1 cervix squamous cell carcinoma, 2 of 2 urothelial carcinomas and 1 of 2 Hodgkin lymphomas are also positive for GATA-3

Tumor type	•	n	GATA3					Total positiv
			0	1+	2+	3+	4+	
Total		99	33	8	7	8	43	66/99 (67%)
ABLE 1. N	ext-generation Immunohi	stochemic	cal Markers D	Discussed in Th	nis Review		Next-gen	eration IHC
ABLE 1 . N Marker	ext-generation Immunohi	stochemia Use	cal Markers D eful in Diagne	iscussed in Th osis of	nis Review		Next-gen "Quali	eration IHC fications"

Bellizzi AM. An Algorithmic Immunohistochemical Approach to Define Tumor Type and Assign Site of Origin. Adv Anat Pathol. 2020;27(3):114-163.

SMH - PITFALLS/POINTS OF ATTENTION



able 4. Proportion of sufficient and optimal results for SMH for the most commonly used RTU IHC systems									
RTU systems	Recomm	nended	Laboratory modified						
	protocol	settings*	protocol settings**						
	Sufficient	Optimal	Sufficient	Optimal					
Dako AS mAb SMMS-1 IR/IS066	100% (9/9)	67% (6/9)	2/3	0/3					
Leica BONB mAb S131 PA0493	2/2	2/2	-	-					
VMS Ultra/XT mAb SMMS-1 760-2704	1/3	0/3	91% (21/23)	57% (13/23)					

* Protocol settings recommended by vendor – Retrieval method and duration, Ab incubation times, detection kit, IHC stainer/equipment.
** Significant modifications: retrieval method, retrieval duration and Ab incubation time altered >25%, detection kit – only protocols
performed on the specified vendor IHC stainer integrated.

able 3. Proportion of a	optimal results for SMH for	the most commonly u	ised antibody as c	oncentrate on the 4
nain IHC systems*				

Concentrated antibodies	Dako Autostainer Link / Classic		Da Om	ako Vo mnis BenchM		tana 'k GX / XT ltra	Leica Bond III / Max	
	TRS pH 9.0	TRS pH 6.1	TRS pH 9.0	TRS pH 6.1	CC1 pH 8.5	CC2 pH 6.0	ER2 pH 9.0	ER1 pH 6.0
mAb SMMS-1	3/10** (30%)	-	1/4	0/1	13/23 (57%)	0/1	2/5 (40%)	-

* Antibody concentration applied as listed above, HIER befores and detection kits used as provided by the vendors of the respective systems.

** (number of optimal results/number of laboratories using this buffer)

Recommended settings with UltraView detection system.

Most common and successful modification: adding UV amplification or use of OptiView.

No RTU for Omnis is available. Four laboratories used the Autostainer RTU on the Omnis unsuccessfully.

Limited data for concentrated formats on Omnis, but possible to achieve an optimal staining.

SMH - ICAPS



SMH - Myosin, smooth muscle heavy chain

Control type	Positive tissue control High expression level	Positive tissue control Low expression levels	Negative tissue control
Tissue	Tonsil	Tonsil	Tonsil
Description	Virtually all smooth muscle cells in vessels must show a moderate to strong cytoplasmic staining reaction.	The vast majority of follicular dendritic cells in the germinal centers must show a weak to moderate distinct cytoplasmic staining reaction.	No staining reaction in epithelial cells. Note, smooth muscle cells in vessels and lamina muscularis will show a moderate to strong staining reaction.
Example	Click to enlarge	Click to enlarge	Click to enlarge



A weak to moderate staining reaction is seen in the follicular dendritic network in the germinal center. A high signal-to-noise ratio is observed.



Optimal SMH staining of the breast ductal carcinoma using same protocol as in Figs. 1a - 4a. A moderate, distinct and continuous staining reaction is seen in the myoepithelial cells lining the breast DCIS component, while the invasive components show no staining.



Fig. 2a Insufficient SMH staining of the tonsil using same protocol as in Fig. 1b. The follicular dendritic network in the germinal center is virtually negative and only vascular smooth muscle cells are demonstrated.



Insufficient SMH staining of the breast ductal carcinoma using same protocol as in Figs. 1a - 4a. No staining is seen in neither the DCIS nor the invasive components and thus not possible to differentiate these two entities.

P63 - PITFALLS/POINTS OF ATTENTION



able 3. Proportion of sufficient and optimal results for p63 for the most commonly used RTU IHC systems								
RTU systems	Recomi protoco	mended I settings*	Laboratory modified protocol settings**					
	Sufficient	Optimal	Sufficient	Optimal				
VMS Ultra/XT mAb 4A4 790-4509	57% (4/7)	0/7	88% (100/114)	52% (59/114)				
Dako AS48 mAb DAK-p63 IR662	91% (11/12)	17% (2/12)	57% (4/7)	0/7				
Dako Omnis mAb DAK-p63 (GA662	85% (17/20)	25% (5/20)	100% (13/13)	62% (8/13)				
Leica Bond mAb 7JUL PA0103	1/4	0/4	0/6	0/6				

* Protocol settings recommended by vendor – Retrieval method and duration, Ab incubation times, detection kit, IHC stainer/equipment.
** Significant modifications: retrieval method, retrieval duration and Ab incubation time altered, detection kit – only protocols performed on the specified vendor IHC stainer integrated.

Table 2. Proportion of optimal results for p63 for the most commonly used antibody as concentrate on the four main IHC systems*

our main Inc systems*										
Concentrated antibodies	Dako/Agilent Autostainer		Dako/Agilent Omnis		Ventana Bench GX / XT	/Roche Mark / Ultra	Leica Bond III / Max			
	TRS pH 9.0	TRS pH 6.1	TRS pH 9.0	TRS pH 6.1	CC1 pH 8.5	CC2 pH 6.0	ER2 pH 9.0	ER1 pH 6.0		
mAb clone 4A4	0/3**	0/1	1/2	-	9/20 (45%)	-	1/7 (14%)	0/1		
mAb clone DAK-p63	0/3	-	4/9 (44%)	0/1	17/24 (71%)	-	0/9	- /		
mAb clone 7JUL	-	-	-	-	0/4	-	0/6	0/1		
A ship of a second s	and Park and Parks	d a base of the	D h C h l	distant for a little state	second and second date	d been the second	dama and the			

* Antibody concentration applied as listed above, HIER buffers and detection kits used as provided by the vendors of the respective systems.

** (number of optimal results/number of laboratories using this buffer).

Vendor recommended protocol based on UltraView and 16-20 min. incubation of primary Ab.

Most common and successful modification was prolonging incubation time and use of OptiView or UltraView with amplification.

Vendor recommended protocol based on HIER in TRS Low pH. Most successful modification was using HIER in TRS High pH.

Less successful performance for 7JUL on the Bond platform. Limited data for Bond users, but conc. 4A4 might

be the best solution.

P63 - PITFALLS/POINTS OF ATTENTION





Courtesy Ole Nielsen, Dept. of Pathology, OUH, Denmark







Left: Strong staining reaction is seen in in squamous epithelium cells in tonsil. Scattered lymphocytes shows a weak, but distinct nuclear staining reaction.



Left: Dispersed cytotrophoblastic cells shows an at least weak to moderate, distinct nuclear staining reaction.



Left: Basal cells in prostate hyperplasia show a moderate to strong nuclear staining reaction.

ECAD - PITFALLS/POINTS OF ATTENTION



Table 1. Antibodies and	d as	sessment marks for ECA	D, run 53	3				
Concentrated antibodies	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	Suff. OPS ²
mAb clone NCH-38	82 1 1	Agilent/Dako Immunologics Thermo S./Neomarkers	57	22	4	1	94%	98%
mAb clone 36	1 1	BD Biosciences Biogenex	0	1	0	1	-	-
mAb clone 36B5	13	Leica/Novocastra	2	10	1	0	92%	100%
mAb clone 4A2C7	4	Life Tech./Invitrogen	2	2	0	0	-	-
mAb clone BS38	1	Nordic Biosite	0	1	0	0	-	-
mAb clone DBM15.49	1	Diagnostic BioSystems	1	0	0	0	-	-
mAb clone ECH-6	2	Zytomed Systems	1	0	1	0	-	-
mAb clone HECD-1	9 1	Life Tech./Invitrogen Takara Bio Inc.	4	5	0	1	90%	100%
mAb clone GM016	1	Genemed	1	0	0	0	-	-
mAb clone SPM471	1	Thermo S./Neomarkers	0	0	1	0	-	-
rmAb EP700Y	5	Cell Marque	0	4	1	0	-	-
rmAb EP6	1	Zeta Corporation	0	1	0	0	-	-
Ready-To-Use antibodies								
mAb clone 36 790-4497	68	Roche/Ventana	54	11	3	0	96%	100%
mAb clone GM016 8229-C010	2	Sakura Finetek	2	0	0	0	100%	5 -
mAb clone NCH-38 GA059	31	Agilent/Dako	31	0	0	0	100%	100%
mAb clone NCH-38 GA059 ³	6	Agilent/Dako	5	1	0	0	-	-
mAb clone NCH-38 IS/IR059	27	Agilent/Dako	26	1	0	0	100%	100%
mAb clone NCH-38 IS/IR059 ³	6	Agilent/Dako	4	2	0	0	-	-
mAb clone MX020 MAB-0738	1	Maixin	0	1	0	0	100%	5 -
mAb clone BS38 MAD-000643QD	1	Master Diagnostica	1	0	0	0	100%	6 -
mAb clone HECD-1 MAD-000761QD	1	Master Diagnostica	1	0	0	0	100%	6 -
mAb clone 35B5 PA0387	6	Leica/Novocastra	0	6	0	0	100%	6 -
rmAb clone EP700Y 760-4440	17	Roche/Ventana	0	2	15	0	13%	-
rmAb clone EP700Y 246R-18	6	Cell Marque	0	1	5	0	-	-
mAb clone EP6 API3012	1	Biocare Medical	0	1	0	0	100%	6 -
Total	298		192	72	31	3	-	
Proportion			65%	24%	10%	1%	89%	

Table 3. Proportion of optimal results for ECAD for the most commonly used antibody as concentrate on the 4 main IHC systems*

Concentrated antibodies	Dako Autostainer Link / Classic		Dako Omnis		Vent BenchMark	tana XT / Ultra	Leica Bond III / Max	
	TRS pH 9.0	TRS pH 6.1	TRS pH 9.0	TRS pH 6.1	CC1 pH 8.5	CC2 pH 6.0	ER2 pH 9.0	ER1 pH 6.0
mAb clone NCH-38	8/10** (80%)	-	1/1	-	32/42 (76%)	-	6/6 (100%)	0/2
* Antibody concord	tration applied	ac licted above	HIED buffore a	and detection I	kite ucod ac prov	ided by the year	adore of the rec	noctivo

* Antibody concentration applied as listed above, HIER buffers and detection kits used as provided by the vendors of the respective systems.

** (number of optimal results/number of laboratories using this buffer)

Concentrated format of mAb NCH-38 works on the main IHC Systems

Table 4. Proportion of sufficient and optimal results for ECAD for the most commonly used RTU IHC systems										
RTU systems	Reco	ommended	Laboratory modified							
	protoc	col settings*	protocol settings**							
	Sufficient	Optimal	Sufficient	Optimal						
Dako AS										
mAb NCH-38	100% (10/10)	100% (10/10)	100% (13/13)	100% (13/13)						
IS/IR059										
Dako Omnis										
mAb NCH-38	100% (21/21)	100% (21/21)	(3/3)	(3/3)						
GA059										
VMS Ultra/XT/GX										
mAb 36	100% (11/11)	72% (8/11)	95% (54/57)	81% (46/57)						
790-4497										

* Protocol settings recommended by vendor - Retrieval method and duration, Ab incubation times, detection kit, IHC stainer/equipment.
** Significant modifications: retrieval method, retrieval duration and Ab incubation time altered >25%, detection kit - only protocols performed on the specified vendor IHC stainer were included.

RTU assays works as "plug-and-play" products. The majority of RTU assays obtain high pass rates – except assays based on rmAb EP700Y

Proportion of sufficient stains (optimal or good).
 Proportion of sufficient stains with extinct states and states and

Proportion of sufficient stains with optimal protocol settings only, see below

 Ready-to-use product developed for a specific semi/fully automated platform by a given manufacturer but inappropriately applied by laboratories on other non-validated semi/fully automatic systems or used manually.

ECAD - ICAPS



ECAD - E-cadherin

Control type	Positive tissue control High expression level	Positive tissue control Low expression levels	Negative tissue control
Tissue	Liver	Liver	Appendix/colon
Description	Virtually all epithelial cells of bile ducts must show a strong membranous staining reaction.	Virtually all hepatocytes must show an at least weak to moderate, distinct membranous staining reaction.	No staining reaction in stromal cells such as lymphocytes, smooth muscle cells and endothelial cells should be seen. Dispersed plasma cells can show a weak membranous staining reaction. Note, epithelial cells should show a strong membranous staining reaction.
Example	Click to enlarge	Click to enlarge	Click to enlarge

Optimal staining result

Too diluted Ab





STILL AWAKE?

KI67 - PITFALLS/POINTS OF ATTENTION



Table 1. Antibodies and assessment marks for Ki67, run B22									
Concentrated antibodies	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	Suff. OPS ²	
mAb clone BS4	1	Nordic Biosite	1	0	0	0	-	-	
mAb clone GM001	1	Genemed	1	0	0	0	-	-	
mAb clone K2	2 1	Zytomed Leica/Novocastra	2	1	0	0	-	-	
mAb clone MIB-1	122 1	Agilent/Dako VWR/Immunologic	72	36	13	2	88%	90%	
mAb clone UMAB107	7	ZSBio	2	4	1	0	86%	80%	
rmAb clone SP6	7 5 3 3 1 1	Thermo/Neomarkers Cell Marque Biocare Spring Bioscience Zytomed Master Diagnostica Diagnostic Biosystems	17	5	1	0	96%	95%	
pAb RB-1510	1	Thermo/Neomarkers	1	0	0	0	-	-	
Ready-To-Use antibodies									
mAb clone GM001 60-0040-7	1	Genemed	1	0	0	0	-	-	
mAb clone K2 PA0230	4	Leica/Novocastra	2	2	0	0	-	-	
mAb clone Ki88 AM370	1	Biogenex	0	1	0	0	-	-	
mAb MIB-1 IR626/IS626	65	Agilent/Dako	34	25	5	1	91%	94%	
mAb MIB-1 GA626	31	Agilent/Dako	25	5	1	0	97%	100%	
mAb clone MIB-1 AM297	1	Biogenex	1	0	0	0	-	-	
mAb clone MM1 PA0118	9	Leica/Novocastra	0	8	1	0	-	-	
mAb clone MX006 MAB-0672	1	Maixin	0	1	0	0	-	-	
rmAb clone SP6 275R	4	Cell Marque	2	1	1	0	-	-	
rmAb clone SP6 PRM 325	1	Biocare	0	1	0	0	-	-	
rmAb clone SP6 MAD-000310QD	1	Master Diagnostica	0	1	0	0			
rmAb clone 30.9 790-4286	131	Roche/Ventana	121	9	1	0	99%	100%	
Total	409		282	100	24	3	-		
Proportion			69%	24%	6%	1%	93%		

1) Proportion of sufficient stains (optimal or good).

2) Proportion of sufficient stains with optimal protocol settings only, see below.



Epitope	Clone	Vendor	Format	Antibody Dilution	Platform	Epitope Retrieval	HIER pH	Visualization System
Ki67	Mib1	Dako	Conc	1:200	Dako Autostainer	HIER	TRS high pH	FLEX+
Ki67	Mib1	Dako	Conc	1:100	Leica Bond	HIER	ER2 (high PH)	Refine
Ki67	Mib1	Dako	Conc	1:200	Ventana Ultra	HIER	CC1 (high pH)	Optiview
Ki67	Mib1	Dako	RTU	_	Dako Autostainer	HIER	TRS low pH	FLEX
Ki67	SP6	CellMarque	Conc	1:75	Dako Autostainer	HIER	TRS low pH	FLEX+, rabbit
Ki67	SP6	CellMarque	Conc	1:75	Leica Bond	HIER	ER2 (high pH)	Refine
Ki67	SP6	CellMarque	Conc	1:150	Ventana Ultra	HIER	CC1 (high pH)	Optiview
Ki67	MM1	Leica	RTU	_	Leica Bond	HIER	ER2 (high pH)	Refine
Ki67	30.9	Ventana	RTU	—	Ventana Ultra	HIER	CC1 (high pH)	Ultraview
PCK	AE1AE3	Dako	Conc	1:100	Dako Autostainer	HIER	TRS High	FLEX+ (mouse)
PCK	AE1AE3	Dako	Conc	1:75	Leica Bond	HIER	ER2 (high pH)	Refine
PCK	AE1AE3	Dako	Conc	1:150	Ventana Ultra	HIER	CC1 (high)	Optiview

Conc indicates concentrated; HIER, heat-induced epitope retrieval; PCK, pan-cytokeratin; RTU, ready-to-use.

KI67 - ICAPS



Ki67 - Ki67

Control type	Positive tissue control High expression level	Positive tissue control Low expression levels	Negative tissue control
Tissue	Tonsil	Tonsil	Liver
Description	Virtually all B-cells in the dark zones of the germinal centres must show a moderate to strong nuclear staining reaction.	The vast majority of B-cells in the light zones of the germinal centres must show a weak to moderate nuclear staining reaction.	No nuclear staining reaction in the hepatocytes should be seen (<1% of hepatocytes should be positive) Note, granulocytes in vessels can show a weak to moderate nuclear staining reaction (at present no explanation for this observation).
Example	Click to enlarge	Click to enlarge	t to enlarge

Optimal protocol settings



NordiC

Too diluted Ab



Breast carc.

ER – PITFALLS / POINTS OF ATTENTION



Table 3. Comparison of pass rates for vendor recommended and laboratory modified RTU protocols										
RTU systems	Vendor rec	ommended	Laboratory modified							
	protocol	settings*	protocol settings**							
	Sufficient	Optimal	Sufficient	Optimal						
Dako AS48		-								
rmAb EP1	5/8 (63%)	2/8 (25%)	15/18 (83%)	8/18 (44%)						
IR084/IS084										
Dako Omnis										
rmAb EP1	26/27 (96%)	14/27 (52%)	22/23 (96%)	15/23 (65%)						
GA084	20/2/ (50/0)									
Leica Bond										
mAb 6F11	0/1	0/1	9/13 (69%)	5/13 (38%)						
PA009/PA0151		-								
VMS Ultra/XT/GX										
rmAb SP1	41/44 (93%)	24/44 (55%)	140/158 (89%)	84/158 (53%)						
790-4324/4325										

* Protocol settings recommended by vendor – Retrieval method and duration, Ab incubation times, detection kit, IHC stainer/equipment.
** Modifications included: retrieval method, retrieval duration, retrieval reagents, Ab incubation time and detection kit. Only protocols performed on the specified vendor IHC stainer are included.

Even with these successful results, changing RTU assays requires internal validation.

For Dako and Ventana products, the most common modification was using a 3-step detection system. For Leica, modification in HIER – changing from low till high pH buffer was made the majority of participants.



Pass rate - RTU vs. LDT - B27-B31



ER – ICAPS



ER - Estrogen Receptor

Control type	Positive tissue c High expression	ontrol level	Positive tissue control Low expression levels	Negative tissue control	Estrogen and	
Tissue	Cervix		Tonsil	Tonsil		
Description	Virtually all squamous epithelial cells, columnar epithelial cells and stromal cells (except lymphocytes and endothelial cells) must show a moderate to strong nuclear staining reaction.		tually all squamous thelial cells, columnar thelial cells and stromal cells cept lymphocytes and dothelial cells) must show a ining reaction. The vast majority of squamous epithelial cells and dispersed follicular dendritic cells / T- cells within the germinal centers must show an at least weak, distinct nuclear staining reaction.		American Society of Clin Kimberly H. Allison, MD ¹ ; M. Elizabeth H. Ha Patrick L. Fitzgibbons, MD ⁶ ; Daniel F. Jane Perlmutter, PhD ¹¹ ; Charles M. W. Fraser Symmans, MD ¹⁰ ; Emina E.	
Abur (big			2010 Recomme	Tracey F. Weisberg,		
Example	Click to enlarge	Optimal i Ongoin Initial a comp	ce. Ongoing quality contro d Initial and ongoing laborassessment should be			
Use o ma eva the slic			SOPs, including routine rials with each batch of lation of internal norma nclusion of normal brea , wherever possible.	rol SOPs should be used t with each batch of to epithelial elements of appropriate control) controls should inclu- with lower percenta		

Progesterone Receptor Testing in Breast Cancer

nical Oncology/College of American Pathologists **Guideline Update**

ammond, MD²; Mitchell Dowsett, PhD³; Shannon E. McKernin⁴; Lisa A. Carey, MD⁵; Hayes, MD⁷; Sunil R. Lakhani, MD^{8,9}; Mariana Chavez-MacGregor, MSc¹⁰; Perou, PhD⁵; Meredith M. Regan, ScD¹²; David L. Rimm, MD, PhD¹³; Torlakovic, MD, PhD^{14,15}; Leticia Varella, MD¹⁶; Giuseppe Viale, MD^{17,18}; MD¹⁹; Lisa M. McShane, PhD²⁰; Antonio C. Wolff, MD²¹

Arch Pathol Lab Med-Vol 144, May 2020

Updated Recommendation

ocedures

ol and equipment maintenance are required.

oratory personnel training and competency e performed.

hat include routine use of external control materials esting and routine evaluation of internal normal or the inclusion of normal breast sections (or other on each tested slide, wherever possible. External de negative and positive samples as well as samples ges of ER expression (such as tonsil). On-slide controls are recommended.







PR – PITFALLS / POINTS OF ATTENTION

able 1. Antibodies and assessment marks for PR, run B31											
Concentrated antibodies	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	OR ²			
nAb clone 16	33 1	Leica Biosystems Monosan	20	12	-	2	94%	59%			
nAb clone cocktail 16 + SAN27	5	Leica Biosystems	2	2	1	-	80%	40%			
mAb clone BP6081	1	Biolynx	-	1	-	-	-	-			
nAb clone PgR 636	13 1	Dako/Agilent Invitrogen	5	4	3	2	64%	36%			
nAb clone PgR 1294	10	Dako/Agilent	8	1	1	-	90%	80%			
nAb clone PR88	1	BioGenex	-	-	-	1	-	-			
mAb clone SP2	1 1	Diagnostic BioSystems Thermo Scientific	2	-	-	-	-	-			
mAb clone SP42	3	Zytomed	-	2	1	-	-	-			
mAb clone YR85	1	Fischer Scientific	-	1	-	-	-	-			
mAb clone ZR4	1	Zeta Corporation	1	-	-	-	-	-			
Ready-To-Use antibodies							Suff.1	OR ²			
nAb clone 16 PA0312 (VRPS³)	6	Leica Biosystems	6	-	-	-	100%	100%			
nAb clone 16 • A0312 (LMPS⁴)	12	Leica Biosystems	10	1	1	-	92%	83%			
nAb clone 16 1AD-000670QD	2	Master Diagnostica	-	-	2	-	-	-			
nAb PgR 636 (R/IS068 (VRPS³)	4	Dako/Agilent	3	1	-	-	-	-			
nAb PgR 636 (R/IS068 (LMPS⁴)	26	Dako/Agilent	21	3	-	2	92%	81%			
nAb PgR 1294 6A090 (VRPS³)	33	Dako/Agilent	10	22	1	-	97%	30%			
nAb PgR 1294 5A090 (LMPS⁴)	20	Dako/Agilent	11	5	4	-	80%	55%			
mAb clone 1E2 790-2223/4296 (VRPS³)	53	Ventana/Roche	44	9	-	-	100%	83%			
mAb clone 1E2 790-2223/4296 (LMPS⁴)	141	Ventana/Roche	108	23	9	1	93%	77%			
nAb clone IHC751 HC751	1	GenomeMe	1	-	-	-	-	-			
mAb clone SP2 Kit-0013	2	Maixin	2	-	-	-	-	-			
mAb clone Y85 3360-C010	4	Sakura Finetek	4	-	-	-	-	-			
nAb PgR 636 PM343	1	Biocare Medical	-	1	-	-	-	-			
otal	377		258	88	23	8					
roportion			68%	23%	6%	2%	92%				

Graph 1. Pass rate in the NordiQC assessments for PR



Table 3. Comparison of pass rates for vendor recommended and laboratory modified RTU protocols										
RTU systems		Vendor reco	mmended		Laboratory modified					
		protocol	settings*		protocol settings**					
		Sufficient	Optimal		Sufficient	Optimal				
Leica BOND MAX/ BOND III										
mAb 16		6/6 (100%)	6/6 (100%)	N	10/10 (100%)	9/10 (90%)				
PA0312										
Dako Autotstainer+/ Autostainer										
Link mAb PgR 636		4/4	3/4		24/25 (96%)	21/25 (84%)				
IS068/IR068										
Dako Omnis										
mAb PgR 1294		32/33 (97%)	10/33 (30%)		16/20 (80%)	11/20 (55%)				
GA090										
Ventana BenchMark GX/XT/Ultra				\mathbf{V}						
rmAb 1E2		53/53 (100%)	44/53 (83%)	1	131/141 (93%)	108/141 (77%)				
790-2223/790-4296										
* Protocol settings recommended by ve	nd	or - Petrieval method a	nd duration Ab neuba	tio	n times detection kit	IHC stainer/equipmer				

Protocol settings recommended by vehiclor Recrieval method and duration, Approcupation times, ** Modifications included: retrieval method, retrieval duration, retrieval reacents, Ab incubation time and detection kit. Only protocols performed on the specified vendor IHC stainer are included

Autostainer RTU: If using Flex+ a pass rate of 95%, 91% optimal.

Omnis RTU: If using Flex+ a pass rate of 92%, 77% optimal.

) Proportion of sufficient results (optimal or good) (≥5 asessed protocols).

) Proportion of optimal results (≥5 asessed protocols).

3) Vendor Recommended Protocol Settings (VRPS) to a specific RTU product applied on the vendor recommended platform(s). 4) Laboratory Modified Protocol Settings (LMPS) to a specific RTU product applied either on the vendor recommended platform(s) or other platforms.



PR – ICAPS



© NordiQC © NordiQC Ositive Nordi

PR - Progesterone Receptor

Control type	Positive tissue control High expression level	Positive tissue control Low expression levels	Negative tissue control
Tissue	Cervix	Cervix	Tonsil
Description	Virtually all columnar epithelial cells most of the stromal cells (except endothelial cells and lymphocytes) must show a moderate to strong nuclear staining reaction.	The vast majority of basal squamous epithelial cells must show an at least weak nuclear staining reaction. Note, PR expression level can be reduced in some samples of uterine cervix e.g. due to post- menopausal status or phase of menstrual cycle.	No and of central importance no nuclear staining reaction in germinal center lymphocytes should be seen.
Example	Click to enlarge	Click to enlarge	Click to enlarge

Optimal protocol settings

rmAb 1E2: Reduced HIER + prolonged inc. time of Ab



HER2 – PITFALLS / POINTS OF ATTENTION



Do not change to a more sensitive detection system!



HER2 – PITFALLS / POINTS OF ATTENTION





HER2 – PITFALLS / POINTS OF ATTENTION



© NordiQC

PD-L1 IC – PITFALLS/POINTS OF ATTENTION

(%)

rate

Pass

CE-IVD / FDA approved PD-L1 assays	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	OR ²
rmAb clone SP142, 741-4860 3	44	Ventana/Roche	30	10	1	3	91%	68%
rmAb clone SP263, 741-4905 ³	6	Ventana/Roche	-	-	6	-	0%	0%
mAb clone 22C3 pharmDX, SK006	3	Dako/Agilent	-	1	-	2	-	-
mAb clone 22C3 pharmDX, GE006	3	Dako/Agilent	-	-	1	2	-	-
Antibodies⁵ for laboratory developed PD-L1 assays, concentrated antibodies	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	OR ²
mAb clone 22C3	4	Dako/Agilent	-	1	-	3	-	-
mAb clone E1L3N	4	Cell Signaling	-	-	4	-	-	-
rmAb clone IHC411	1	GenomeMe	-	-	1	-	-	-
rmAb clone CAL10	1 1	Zytomed Biocare Medical	-	-	2	-	-	-
rmAb clone ZR3	2	Zeta Corporation	-	-	-	2	-	-
Ready-To-Use antibodies ⁶	n	Vendor	Optimal	Good	Borderline	Poor	Suff.1	OR ²
rmAb clone SP142, 790-4860 (VRPS)⁴	14	Ventana/Roche	10	3	1	-	93%	71%
rmAb clone SP142, 790-4860 (LMPS)⁵	36	Ventana/Roche	25	6	2	3	86%	69%
rmAb clone SP263, 790-4905	1	Ventana/Roche	-	-	1	-	-	-
rmAb clone 73-10, PA0832	1	Leica Biosystems	-	-	1	-	-	-
rmAb clone MXR003, RMA-0732	2	Maixin	-	-	2	-	-	-
rmAb clone ZR3, GT228002	1	Gene Tech	-	-	-	1	-	-
mAb clone 405-9A11, PDM572	1	Diagnostic BioSystems	-	-	-	1	-	-
Total	125		65	21	22	17		
Proportion			52%	17%	18%	13%	69%	

Table 2. Assessment marks for IHC assays and antibodies run C9. PD-L1 IC (TECENTRIO®)

Proportion of sufficient stains (optimal or good) (≥5 assessed protocols).

Proportion of optimal results (≥5 assessed protocols).

This product has a locked protocol on all BenchMark platforms and cannot be changed.

4) Vendor recommended protocol settings - RTU product used in compliance to protocol settings, platform and package insert.

5) Laboratory modified protocol settings for a RTU product applied either on the vendor recommended platform(s) or other platforms.

6) mAb: mouse monoclonal antibody, rmAb: rabbit monoclonal antibody.

7) Ready-To-Use antibodies without predictive claim.



Characteristics of insufficient results in the NordiQC PD-L1 IC assessments.













rmAb SP142

mAb 22C3

rmAb SP142

rmAb ZR3



CONGRATULATIONS!

...YOU SURVIVED! THANK YOU FOR YOUR ATTENTION

BONUS – P120



No NordiQC data available for p120 Catenin. For the p120 stains below, a concentrated format of the mAb clone MRQ-5 is used.



E-CAD: membranous staining reaction in (most) ductal breast carcinomas, negative in (most) lobular breast carcinomas.

p120: membranous staining reaction in (most) ductal breast carcinomas, cytoplasmic staining reaction in (most) lobular breast carcinomas.

Ductal breast carcinoma

Lobular breast carcinoma

BONUS – P120 ICAPS



No NordiQC data available for p120 Catenin. For the p120 stains below, a concentrated format of the mAb clone MRQ-5 is used.

Liver: Hepatocytes must show a weak to moderate membranous staining



Appendix: Columnar epithelial cells must show a strong membranous staining.