

Recommended protocol for VIM

Obtained in run 52

27 Dec 2017

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: SP20
Producer: NeoMarkers
Product no. / lot no.: RM-9120-S / 9120S1702J
Diluent: Normal antibody Diluent
Diluent producer / no.: Immunologic / VWRKBD09-999
Dilution factor: 1:200
Incubation time / temperature: 20 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana CC1
Heating time at max. temp.: 64 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Ventana
Product / no: UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer: 8 min.
Amplifier: Ultraview/iView Amplification Kit
Incubation time amplifier: 8 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no: ultraView Universal DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 36°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.