Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for OCT3/4

Obtained in run 53

13 Feb 2018

Immunostainer

Type:	Manual

Primary antibody

Clone:	NRG1.1
Producer:	Diagnostic BioSystems
Product no. / lot no.:	Mob 464-05 / W053
Diluent:	Diamond - Antibody Diluent
Diluent producer / no.:	Cell Marque / 938B
Dilution factor:	1:50
Incubation time / temperature:	900 min. / 4°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Dako TRS High pH (3-1)
Heating time at max. temp.:	20 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Dako
Product / no:	EnVision FLEX+ / K8012/DM822
Linker:	Linker, Mouse
Incubation time linker:	10 min.
Incubation time polymer:	10 min.
Incubation temperature:	37°C

Chromogen

Producer:	Dako
Product / no:	DAB+ / DM827 / DM823 / K800
Incubation time / temperature:	5 min. / 37°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.