Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for OCT3/4

Obtained in run 53

20 Feb 2018

Immunostainer

Type:	Dako Autostainer

Primary antibody

Clone:	C10
Producer:	Santa Cruz
Product no. / lot no.:	SC-5279 / A2315
Diluent:	Da Vinci green
Diluent producer / no.:	Biocare / PD900M
Dilution factor:	1:75
Incubation time / temperature:	30 min. / 22°C

Epitope retrieval, HIER

Device:	Microwave oven
Buffer:	Dako TRS High pH
Heating time at max. temp.:	10 min.
Maximum heating temp.:	100°C

Visualization system

Producer:	Biocae
Product / no:	MACH4 / M4U534L
Linker:	MACH4 mouse probe
Incubation time linker:	10 min.
Incubation time polymer:	15 min.
Incubation temperature:	22°C

Chromogen

Producer:	Biocare
Product / no:	Betazoid DAB / BDB2004
Incubation time / temperature:	5 min. / 22°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.