Nordic Immunohistochemical Quality Control
Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CK-PAN

Obtained in run 54

30 Aug 2018

Immunostainer

Type:	Thermo Autostainer 36/48/72

Primary antibody

Clone:	BS5
Producer:	Nordic Biosite
Product no. / lot no.:	BSH-7124-100 / BSH1-10t
Diluent:	Antibody Diluent
Dilution factor:	1:100
Incubation time / temperature:	30 min. / 21°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.:	20 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Nordic Biosite
Product / no:	BioSite Histo Plus HRP Polymer anti-Mouse / KDB-10007
Linker:	None
Incubation time polymer:	30 min.
Incubation temperature:	21°C

Chromogen

Producer:	Nordic Biosite
Product / no:	High contrast DAB / BCB-20032
Incubation time / temperature:	10 min. / 21°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.