**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for WT1

Obtained in run 55

12 Dec 2018

**Immunostainer**

| | |
|---|---|
| Type: | Dako Omnis |

**Primary antibody**

| | |
|---|---|
| Clone: | 6F-H2 |
| Producer: | Dako/Agilent |
| Product no. / lot no.: | M3561 / 10130603 |
| Diluent: | Antibody Diluent |
| Diluent producer / no.: | Dako / K8006 |
| Dilution factor: | 1:100 |
| Incubation time / temperature: | 20 min. / 32°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Dako Omnis Target Retrieval Solution, High pH |
| Heating time at max. temp.: | 30 min. |
| Maximum heating temp.: | 97°C |

**Visualization system**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | EnVision Flex / GV800/GV823 |
| Linker: | Mouse LINKER |
| Incubation time linker: | 10 min. |
| Incubation time polymer: | 20 min. |
| Incubation temperature: | 32°C |

**Chromogen**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | DAB+ Substrate Chromogen System / GV825 |
| Incubation time / temperature: | 5 min. / 32°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.