

## Recommended protocol for IgM

Obtained in run

20 Sep 2015

**Immunostainer**

Type: Ventana Benchmark Ultra

**Primary antibody**

Clone: Polyclonal  
Producer: Dako  
Product no. / lot no.: A0425 / 42162  
Diluent: Antibody Diluent with Casein  
Diluent producer / no.: Ventana / 760-219  
Dilution factor: 1:300  
Incubation time / temperature: 24 min. / 37°C

**Epitope retrieval, HIER**

Device: On Board / On Machine  
Buffer: Ventana Ultra CC1  
Heating time at max. temp.: 30 min.  
Maximum heating temp.: 95°C

**Visualization system**

Producer: Ventana  
Product / no: UltraView Universal DAB Detection Kit / 760-500  
Incubation time polymer: 8 min.  
Incubation temperature: 36°C

**Chromogen**

Producer: Ventana  
Product / no: ultraView Universa DAB Detection Kit / 760-500  
Incubation time / temperature: 8 min. / °C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.