

Recommended protocol for Bcl-2

Obtained in run 57

21 Aug 2019

Immunostainer

Type: intelliPATH

Primary antibody

Clone: 100/D5
Producer: Biocare Medical
Product no. / lot no.: PM003 / 080818
Diluent: None
Dilution factor: 1:1
Incubation time / temperature: 45 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Biocare Medical Reveal Decloaker
Heating time at max. temp.: 15 min.
Maximum heating temp.: 110°C

Visualization system

Producer: Biocare Medical
Product / no: MACH 4 HRP Polymer / MRH534
Linker: MACH 4 Mouse Probe
Incubation time linker: 10 min.
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Biocare Medical
Product / no: intelliPATH DAB chromogen / IPK5010
Incubation time / temperature: 5 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.