

Recommended protocol for CK-PAN

Obtained in run 58

09 Dec 2019

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: AE1/AE3/PCK26
Producer: Ventana/Roche
Product no. / lot no.: 760-2135 / F07799
Format: Ready-To-Use (prediluted)
Incubation time / temperature: 8 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana CC1
Heating time at max. temp.: 36 min.
Maximum heating temp.: 95°C

Epitope retrieval, proteolysis

Enzyme: Protease 3
Enzyme producer / no: Ventana / 760-2020
Incubation time / temp: 4 min. / 36°C

Visualization system

Producer: Ventana
Product / no: UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer: 8 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no: ultraView Universal DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 36°C
Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.