

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CK-PAN

Obtained in run 58

09 Dec 2019

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: AE1/AE3/PCK26
Producer: Ventana/Roche

Product no. / lot no.: 760-2135 / F07799

Format: Ready-To-Use (prediluted)

Incubation time / temperature: 8 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Ventana CC1

Heating time at max. temp.: 36 min. Maximum heating temp.: 95°C

Epitope retrieval, proteolysis

Enzyme: Protease 3

Enzyme producer / no: Ventana / 760-2020

Incubation time / temp: 4 min. / 36°C

Visualization system

Producer: Ventana

Product / no: UltraView Universal DAB Detection Kit / 760-500

Incubation time polymer: 8 min. Incubation temperature: 36°C

Chromogen

Producer: Ventana

Product / no: ultraView Universal DAB Detection Kit / 760-500

Incubation time / temperature: 8 min. / 36°C

Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.