Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD45

Obtained in run 59

19 Feb 2020

Immunostainer

Type:	Tissue-Tek Genie Advanced Stai

Primary antibody

Clone:	GR009
Producer:	Sakura Finetek
Product no. / lot no.:	8271-C010 / 8271C30218
Diluent:	None
Diluent producer / no.:	None / None
Dilution factor:	1:0
Incubation time / temperature:	30 min. / 20°C

Epitope retrieval, HIER

Device:	Other
Buffer:	Sakura Finetek Tissue-Tek Genie High pH Antigen Retriev
Heating time at max. temp.:	60 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Sakura Finetek
Product / no:	Tissue-Tek Genie Pro Detection Kit, DAB / 8826-K250
Linker:	Tissue-Tek Genie Pro Detection Kit, DAB
Incubation time linker:	15 min.
Incubation time polymer:	20 min.
Incubation temperature:	20°C

Chromogen

Producer:	Sakura Finetek
Product / no:	Tissue-Tek Genie Pro Detection Kit, DAB / 8826-K250
Incubation time / temperature:	6 min. / 25°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.