

## Recommended protocol for Ki67

Obtained in run B22

08 Sep 2016

### Immunostainer

Type: Manual

### Primary antibody

Clone: K2  
Producer: Zytomed Systems GmbH  
Product no. / lot no.: MSK018 / Q849  
Diluent: Antibody Diluent  
Dilution factor: 1:300  
Incubation time / temperature: 60 min. / 22°C

### Epitope retrieval, HIER

Device: Pressure Cooker  
Buffer: Std Citrate pH 6  
Heating time at max. temp.: 8 min.  
Maximum heating temp.: 110°C

### Visualization system

Producer: Zytomed Systems  
Product / no: HRP-Polymer / POLHRP100  
Linker: Postblock  
Incubation time linker: 20 min.  
Incubation time polymer: 30 min.  
Incubation temperature: 22°C

### Chromogen

Producer: Zytomed Systems  
Product / no: DAB Substrat Kit / DAB500plus  
Incubation time / temperature: 10 min. / 22°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.