

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## **Recommended protocol for SOX10**

## Obtained in run 60

31 Aug 2020

**Immunostainer** 

Type: Ventana Benchmark Ultra

**Primary antibody** 

Clone: SP267

Producer: Ventana/Roche

Product no. / lot no.: 760-4968 / V0002036

Format: Ready-To-Use (prediluted)

Incubation time / temperature: 32 min. / 23°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Ventana Ultra CC1

Heating time at max. temp.: 64 min. Maximum heating temp.: 95°C

Visualization system

Producer: Ventana

Product / no: UltraView Universal DAB Detection Kit / 760-500

Incubation time polymer: 8 min. Incubation temperature: 23°C

Chromogen

Producer: Ventana

Product / no: ultraView Universal DAB Detection Kit / 760-500

Incubation time / temperature: 4 min. / 23°C

Enhancement: CuSO4

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.