

Recommended protocol for Ki67

Obtained in run B22

12 Sep 2016

Immunostainer

Type: Manual

Primary antibody

Clone: UMAB107
Producer: ZSBio
Product no. / lot no.: ZM-0166 / V010107
Diluent: antibody diluent
Dilution factor: 1:200
Incubation time / temperature: 90 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: ZSBio Tris/EDTA pH8.0
Heating time at max. temp.: 3 min.
Maximum heating temp.: 121°C

Visualization system

Producer: ZSBio
Product / no: iVision Rabbit/Mouse Detection Kit / PV-8000
Linker: None
Incubation time polymer: 25 min.
Incubation temperature: 25°C

Chromogen

Producer: ZSBio
Product / no: DAB kit / ZLI-9018
Incubation time / temperature: 2 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.