

## Recommended protocol for PAX8

Obtained in run 62

17 Feb 2021

**Immunostainer**

Type: Intellipath FLX

**Primary antibody**

Clone: BC12  
Producer: Biocare Medical  
Product no. / lot no.: API438 / 120120A  
Diluent: None  
Diluent producer / no.: None / None  
Dilution factor: 1:1  
Incubation time / temperature: 30 min. / 25°C

**Epitope retrieval, HIER**

Device: Pressure Cooker  
Buffer: Biocare Medical Diva Decloker  
Heating time at max. temp.: 15 min.  
Maximum heating temp.: 110°C

**Visualization system**

Producer: Biocare Medical  
Product / no: MACH 4 HRP Polymer / MHR534  
Linker: MACH 4 Mouse Probe  
Incubation time linker: 10 min.  
Incubation time polymer: 20 min.  
Incubation temperature: 25°C

**Chromogen**

Producer: Biocare Medical  
Product / no: intelliPATH DAB / IPK5010  
Incubation time / temperature: 5 min. / 25°C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.