

## Recommended protocol for CK7

Obtained in run 62

13 Apr 2021

**Immunostainer**

Type: Thermo Autostainer 36/48/72

**Primary antibody**

Clone: OV-TL 12/30  
Producer: ThermoScientific  
Product no. / lot no.: MS-1352-P / 1352P1912B  
Diluent: Normal Antibody Diluent  
Diluent producer / no.: Immunologic / VWRKBD09-999  
Dilution factor: 1:200  
Incubation time / temperature: 60 min. / 25°C

**Epitope retrieval, HIER**

Device: PT-link / PT-module  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 99°C

**Visualization system**

Producer: Immunologic  
Product / no: Mouse/Rabbit HRP / VWRKDPVB999HRP  
Linker: None  
Incubation time polymer: 30 min.  
Incubation temperature: 25°C

**Chromogen**

Producer: Scytek  
Product / no: Dab chromogen / ACB060  
Incubation time / temperature: 8 min. / 25°C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.