

## Recommended protocol for CK20

Obtained in run 62

25 Feb 2021

**Immunostainer**

Type: Manual

**Primary antibody**

Clone: ZM42  
Producer: Zeta Corporation  
Product no. / lot no.: Z2349 / 1806083  
Diluent: Normal Antibody Diluent  
Diluent producer / no.: Zeta Corporation / ZD19  
Dilution factor: 1:200  
Incubation time / temperature: 45 min. / 25°C

**Epitope retrieval, HIER**

Device: Pressure Cooker  
Buffer: Zeta Corporation Tris-EDTA HIER Solution pH 9.0  
Heating time at max. temp.: 45 min.  
Maximum heating temp.: 95°C

**Visualization system**

Producer: Zeta Corporation  
Product / no: Zeta Universal Polymer / ZD8  
Linker: None  
Incubation time polymer: 30 min.  
Incubation temperature: 25°C

**Chromogen**

Producer: Zeta Corporation  
Product / no: DAB Chromogen / ZD11  
Incubation time / temperature: 5 min. / 25°C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.