Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CK7

Obtained in run 62

08 Mar 2021

Immunostainer

Type:	Thermo Autostainer 36/48/72

Primary antibody

Clone:	OV-TL12/30
Producer:	VITRO SA
Product no. / lot no.:	MAD-001004QD / 10040031
Diluent:	None
Diluent producer / no.:	None / None
Dilution factor:	1:1
Incubation time / temperature:	10 min. / 21°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Std EDTA / EGTA pH 8
Heating time at max. temp.:	20 min.
Maximum heating temp.:	95°C

Visualization system

Producer:	VITRO SA
Product / no:	MASTER POLYMER PLUS / MAD-000237QD
Linker:	None
Incubation time polymer:	15 min.
Incubation temperature:	21°C

Chromogen

Producer:	VITRO SA
Product / no:	DAB / MAD-001816QK
Incubation time / temperature:	5 min. / 21°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.