

## Recommended protocol for CD79a

Obtained in run 63

09 Aug 2021

**Immunostainer**

Type: CNT330

**Primary antibody**

Clone: C4C4  
Producer: Celnovte  
Product no. / lot no.: CCM-0734 / 202106201  
Diluent: None  
Diluent producer / no.: None / None  
Dilution factor: 1:0  
Incubation time / temperature: 30 min. / 25°C

**Epitope retrieval, HIER**

Device: On Board / On Machine  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 100°C

**Visualization system**

Producer: Celnovte  
Product / no: CNTVision Super PolyDetection / SD5100  
Linker: CNTVision Super PolyDetection  
Incubation time linker: 20 min.  
Incubation time polymer: 20 min.  
Incubation temperature: 25°C

**Chromogen**

Producer: Celnovte  
Product / no: CNTVision Super PolyDetection / SD5100  
Incubation time / temperature: 6 min. / 25°C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.