Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD45

Obtained in run

27 Sep 2015

Immunostainer

Type:	ImmunoVision, Autostainer

Primary antibody

Clone:	2B11+PD7/26
Producer:	Dako
Product no. / lot no.:	M0701 / 72600
Diluent:	Antibody Diluent
Diluent producer / no.:	Immunologic / NS
Dilution factor:	1:200
Incubation time / temperature:	45 min. / 20°C

Epitope retrieval, HIER

Device:	Water bath
Buffer:	Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.:	60 min.
Maximum heating temp.:	90°C

Visualization system

Producer:	BrigthVision
Product / no:	NS / DPVM-500HRP
Incubation time polymer:	25 min.
Incubation temperature:	20°C

Chromogen

Producer:	Dako
Product / no:	DAB+ / K3467 / K3468
Incubation time / temperature:	5 min. / 20°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.