Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for AMACR

Obtained in run 65

17 Feb 2022

Immunostainer

Type:	DBS Montage 360 system

Primary antibody

Clone:	13H4
Producer:	DBS
Product no. / lot no.:	RMAB078 / R713
Diluent:	Primary antibody diluent
Diluent producer / no.:	DBS / K004
Dilution factor:	1:100
Incubation time / temperature:	30 min. / 20°C

Epitope retrieval, HIER

Device:	Pressure Cooker
Buffer:	DBS Diagnostic Biosystems EDTA Buffer, pH8.0
Heating time at max. temp.:	15 min.
Maximum heating temp.:	105°C

Visualization system

Producer:	DBS
Product / no:	PolyVue Plus Mouse/Rabbit HRP/DAB detection system /
Linker:	PolyVue Plus Mouse/Rabbit HRP Enhancer
Incubation time linker:	10 min.
Incubation time polymer:	10 min.
Incubation temperature:	20°C

Chromogen

Producer:	DBS
Product / no:	Stable DAB/Plus / K047
Incubation time / temperature:	5 min. / 20°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.