Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for Napsin A

Obtained in run 66

22 Aug 2022

Immunostainer

Type:	Intellipath Biocare MEdical

Primary antibody

Clone:	BC15
Producer:	Biocare Medical
Product no. / lot no.:	API3043AA / 011222A
Diluent:	Da Vinci Green
Diluent producer / no.:	Biocare Medical / PD900L
Dilution factor:	1:1
Incubation time / temperature:	30 min. / 25°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Dako TRS High pH (3-1)
Heating time at max. temp.:	20 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Biocare MEdical
Product / no:	MACH-4UNIVERSAL HRP-POLYMER DETECTION SYST
Linker:	MACH4 Mouse Probe
Incubation time linker:	10 min.
Incubation time polymer:	20 min.
Amplifier:	MACH4 HRP Polymer
Incubation time amplifier:	NA min.
Incubation temperature:	25°C

Chromogen

Producer:	Biocare Medical
Product / no:	Betazoid DAB chromogen / BDB900F
Incubation time / temperature:	5 min. / 25°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided

information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.