

Recommended protocol for CD4

Obtained in run 67

04 Jan 2023

Immunostainer

Type: Thermo Autostainer 36/48/72

Primary antibody

Clone: EP204
Producer: Vitro-Master Diagnostica
Product no. / lot no.: MAD-000600QD / 06000022S
Diluent: Antibody Diluent
Diluent producer / no.: Vitro-Master Diagnos / MAD-004048R
Dilution factor: 1:0
Incubation time / temperature: 10 min. / 25°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Std EDTA / EGTA pH 8
Heating time at max. temp.: 20 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Vitro-Master Diagnos
Product / no: Master Polymer Plus Detection System (Peroxidase) / MA
Linker: Primary Antibodies Amplifier Master
Incubation time linker: 10 min.
Incubation time polymer: 15 min.
Amplifier: Master Polymer Plus HRP
Incubation time amplifier: NA min.
Incubation temperature: 25°C

Chromogen

Producer: Vitro-Master Diagnos
Product / no: DAB Chromogen Concentrate / MAD-001811QK-B
Incubation time / temperature: 5 min. / 25°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided



Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.