

Recommended protocol for MSH2

Obtained in run 68

20 Apr 2023

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: RED2
Producer: Beijing zhongshan
Product no. / lot no.: ZA-0622 / 22062110
Diluent: Antibody dilunet
Diluent producer / no.: Beijing Zhongshan / ZLI-9030
Dilution factor: 1:300
Incubation time / temperature: 16 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana Ultra CC1
Heating time at max. temp.: 64 min.
Maximum heating temp.: 99°C

Visualization system

Producer: Ventana
Product / no: OptiView DAB IHC Detection Kit / 760-700
Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no: ultraView Universal DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 36°C
Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.