

Recommended protocol for MSH2

Obtained in run 68

02 Mar 2023

Immunostainer

Type: LYNX480

Primary antibody

Clone: BPM6143
Producer: Biolynx Biotechnology Co., Ltd
Product no. / lot no.: I11372E / B230221043
Diluent: PBS with sucrose (pH7.2-7.4)
Diluent producer / no.: Biolynx / I30012B
Dilution factor: 1:10
Incubation time / temperature: 30 min. / 25°C

Epitope retrieval, HIER

Device: Other
Buffer: Biolynx Antigen Retrieval 2 (EDTA)
Heating time at max. temp.: 20 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Biolynx
Product / no: BXV Visualization System / I20032C
Linker: Post Primary Antibody
Incubation time linker: 10 min.
Incubation time polymer: 10 min.
Incubation temperature: 25°C

Chromogen

Producer: Biolynx
Product / no: DAB / I20032C
Incubation time / temperature: 10 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.