

Recommended protocol for MSH2

Obtained in run 68

08 Mar 2023

Immunostainer

Type: Manual

Primary antibody

Clone: H3Y7
Producer: DaTe Bioengineering Technology
Product no. / lot no.: DTBL0216901 / 230101471
Diluent: None
Diluent producer / no.: None / None
Dilution factor: 1:0
Incubation time / temperature: 60 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: DaTe Bioengineering Epitope Retrieval Solution(EDTA)
Heating time at max. temp.: 2min30s min.
Maximum heating temp.: 100°C

Visualization system

Producer: DaTe Bioengineering
Product / no: IHC Detection Kit / DTBL0500301
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 25°C

Chromogen

Producer: DaTe Bioengineering
Product / no: DAB (IHC Detection Kit) / DTBL0500301
Incubation time / temperature: 5 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.