

## Recommended protocol for URO II/III

Obtained in run 68

13 Mar 2023

**Immunostainer**

Type: AUTOSTAINER LINK 480

**Primary antibody**

Clone: UROPLAKIN II (BC21)  
Producer: MASTER DIAGNOSTICA  
Product no. / lot no.: MAD-000773QD-7/N / 07730013  
Diluent: None  
Diluent producer / no.: None / None  
Dilution factor: 1:0  
Incubation time / temperature: 20 min. / 25°C

**Epitope retrieval, HIER**

Device: PT-link / PT-module  
Buffer: MASTER DIAGNOSTICA EDTA BUFFER 10X PH8  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 95°C

**Visualization system**

Producer: MASTER DIAGNOSTICA  
Product / no: MASTER POLYMER PLUS DETECTION SYSTEM / 2370  
Linker: None  
Incubation time polymer: 15 min.  
Amplifier: PRIMARI ANTIBODIES AMPLIFIER MASTER  
Incubation time amplifier: NA min.  
Incubation temperature: 25°C

**Chromogen**

Producer: MASTER DIAGNOSTICA  
Product / no: DAB ENHANCER / MAD-001811QK-B  
Incubation time / temperature: 5 min. / 25°C  
Enhancement: CuSO<sub>4</sub>

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect



## **Nordic Immunohistochemical Quality Control**

Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

the most recent developments in immunohistochemistry at any point in time.