Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for URO II/III

Obtained in run 68

13 Mar 2023

Immunostainer

Type:	AUTOSTAINER LINK 480

Primary antibody

Clone:	UROPLAKIN II (BC21)
Producer:	MASTER DIAGNOSTICA
Product no. / lot no.:	MAD-000773QD-7/N / 07730013
Diluent:	None
Diluent producer / no.:	None / None
Dilution factor:	1:0
Incubation time / temperature:	20 min. / 25°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	MASTER DIAGNOSTICA EDTA BUFFER 10X PH8
Heating time at max. temp.:	20 min.
Maximum heating temp.:	95°C

Visualization system

Producer:	MASTER DIAGNÃ“STICA
Product / no:	MASTER POLYMER PLUS DETECTION SYSTEM / 237(
Linker:	None
Incubation time polymer:	15 min.
Amplifier:	PRIMARI ANTIBODIES AMPLIFIER MASTER
Incubation time amplifier:	NA min.
Incubation temperature:	25°C

Chromogen

Producer:	MASTER DIAGNÃ“STICA
Product / no:	DAB ENHANCER / MAD-001811QK-B
Incubation time / temperature:	5 min. / 25°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect

the most recent developments in immunohistochemistry at any point in time.