**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for CD138

Obtained in run 69

13 Jul 2023

**Immunostainer**

| | |
|---|---|
| Type: | Leica BOND Max |

**Primary antibody**

| | |
|---|---|
| Clone: | IHC138 |
| Producer: | GenomeMe |
| Product no. / lot no.: | IHC138 / C13832032 |
| Diluent: | GenomeMe Antibody Diluent |
| Diluent producer / no.: | GenomeMe / IHC000 |
| Dilution factor: | 1:200 |
| Incubation time / temperature: | 15 min. / 20°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Leica Bond Epitope Retrieval Solution 2 |
| Heating time at max. temp.: | 20 min. |
| Maximum heating temp.: | 100°C |

**Visualization system**

| | |
|---|---|
| Producer: | Leica |
| Product / no: | Bond Refine / DS9800 |
| Incubation time linker: | 8 min. |
| Incubation time polymer: | 8 min. |
| Incubation temperature: | 20°C |

**Chromogen**

| | |
|---|---|
| Producer: | Leica |
| Product / no: | Bond Refine / DS9800 |
| Incubation time / temperature: | 10 min. / 20°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.