**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for PSA

Obtained in run 69

18 Jul 2023

**Immunostainer**

| | |
|---|---|
| Type: | Ventana Benchmark Ultra |

**Primary antibody**

| | |
|---|---|
| Clone: | Polyclonal |
| Producer: | Dako |
| Product no. / lot no.: | A0562 / 2005991 |
| Diluent: | Antibody Diluent |
| Diluent producer / no.: | Ventana / 251-018 |
| Dilution factor: | 1:3000 |
| Incubation time / temperature: | 32 min. / 36°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Ventana Ultra CC1 |
| Heating time at max. temp.: | 36 min. |
| Maximum heating temp.: | 95°C |

**Visualization system**

| | |
|---|---|
| Producer: | Ventana |
| Product / no: | UltraView Universal DAB Detection Kit / 760-500 |
| Incubation time polymer: | 8 min. |
| Incubation temperature: | 36°C |

**Chromogen**

| | |
|---|---|
| Producer: | Ventana |
| Product / no: | ultraView Universal DAB Detection Kit / 760-500 |
| Incubation time / temperature: | 8 min. / 36°C |
| Enhancement: | CuSO4 |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.