Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD138

Obtained in run 69

24 Aug 2023

Immunostainer

Type:	Gene Tech GeneStainer

Primary antibody

Clone:	GR106
Producer:	Gene Tech
Product no. / lot no.:	GT245102 / 2023081501
Diluent:	None
Diluent producer / no.:	None / None
Dilution factor:	1:0
Incubation time / temperature:	50 min. / 25°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.:	20 min.
Maximum heating temp.:	95°C

Visualization system

Producer:	Gene Tech
Product / no:	GTVision / GK8007
Linker:	None
Incubation time polymer:	20 min.
Incubation temperature:	25°C

Chromogen

Producer:	Gene Tech
Product / no:	DAB / GK8007
Incubation time / temperature:	5 min. / 25°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.