

## Recommended protocol for CD138

Obtained in run 69

24 Aug 2023

**Immunostainer**

Type: Gene Tech GeneStainer

**Primary antibody**

Clone: GR106  
Producer: Gene Tech  
Product no. / lot no.: GT245102 / 2023081501  
Diluent: None  
Diluent producer / no.: None / None  
Dilution factor: 1:0  
Incubation time / temperature: 50 min. / 25°C

**Epitope retrieval, HIER**

Device: On Board / On Machine  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 95°C

**Visualization system**

Producer: Gene Tech  
Product / no: GTVision / GK8007  
Linker: None  
Incubation time polymer: 20 min.  
Incubation temperature: 25°C

**Chromogen**

Producer: Gene Tech  
Product / no: DAB / GK8007  
Incubation time / temperature: 5 min. / 25°C  
Enhancement: None

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.