**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for CD5

Obtained in run 69

29 Aug 2023

**Immunostainer**

| | |
|---|---|
| Type: | Dako Omnis |

**Primary antibody**

| | |
|---|---|
| Clone: | SP19 |
| Producer: | Zytomed |
| Product no. / lot no.: | SP19 / A8170 |
| Diluent: | Antibody Diluent |
| Diluent producer / no.: | Dako / K8006 |
| Dilution factor: | 1:40 |
| Incubation time / temperature: | 20 min. / 36°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Dako Omnis Target Retrieval Solution, High pH |
| Heating time at max. temp.: | 30 min. |
| Maximum heating temp.: | 98°C |

**Visualization system**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | EnVision Flex / GV800/GV823 |
| Linker: | None |
| Incubation time polymer: | 20 min. |
| Incubation temperature: | 36°C |

**Chromogen**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | DAB+ Substrate Chromogen System / GV825 |
| Incubation time / temperature: | 5 min. / 21°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.