Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD34

Obtained in run

06 Jun 2016

Immunostainer

Type:	IntelliPath FLX, Biocare

Primary antibody

Clone:	QBend10
Producer:	Dako
Product no. / lot no.:	M7165 / 20023391
Diluent:	Da Vinci green
Diluent producer / no.:	Biocare / NS
Dilution factor:	1:100
Incubation time / temperature:	30 min. / 20°C

Epitope retrieval, HIER

Device:	Pressure Cooker
Buffer:	Biocare Diva Decloaker pH 6,2
Heating time at max. temp.:	15 min.
Maximum heating temp.:	110°C

Visualization system

Producer:	Biocare
Product / no:	MACH 4 / NS
Linker:	MACH 4
Incubation time linker:	10 min.
Incubation time polymer:	20 min.
Incubation temperature:	20°C

Chromogen

Producer:	Biocare
Product / no:	DAB MACH 4 / NS
Incubation time / temperature:	5 min. / 20°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.