**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for Bcl-6

Obtained in run 70

15 Dec 2023

**Immunostainer**

| | |
|---|---|
| Type: | Dako Omnis |

**Primary antibody**

| | |
|---|---|
| Clone: | PG-B6p |
| Producer: | Dako/Agilent |
| Product no. / lot no.: | GA625 / 41471404 |
| Format: | Ready-To-Use (prediluted) |
| Incubation time / temperature: | 20 min. / 32°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Dako Omnis Target Retrieval Solution, High pH |
| Heating time at max. temp.: | 10 min. |
| Maximum heating temp.: | 97°C |

**Visualization system**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | EnVision Flex / GV800/GV823 |
| Linker: | Mouse LINKER |
| Incubation time linker: | 10 min. |
| Incubation time polymer: | 20 min. |
| Incubation temperature: | 32°C |

**Chromogen**

| | |
|---|---|
| Producer: | Dako Omnis |
| Product / no: | DAB+ Substrate Chromogen System / GV825 |
| Incubation time / temperature: | 5 min. / 32°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.