Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for Bcl-6

Obtained in run 70

10 Jan 2024

Immunostainer

Type:	Dartmon - AS330PLUS

Primary antibody

Clone:	DA005
Producer:	Dartmon Biotechnology Co., Ltd
Product no. / lot no.:	RMB1A051 / 0512308301
Diluent:	None
Diluent producer / no.:	None / None
Dilution factor:	1:0
Incubation time / temperature:	30 min. / 25°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Dartmon Immunohistochemical Antigen Retrieval Buffer
Heating time at max. temp.:	15 min.
Maximum heating temp.:	100°C

Visualization system

Producer:	Dartmon
Product / no:	Immunochromogenic Reagent / DMRD4044
Linker:	lmmunochromogenic Reagent
Incubation time linker:	20 min.
Incubation time polymer:	20 min.
Incubation temperature:	25°C

Chromogen

Producer:	Dartmon
Product / no:	Immunochromogenic Reagent / DMRD4044
Incubation time / temperature:	6 min. / 25°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.