Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD20

Obtained in run 71

15 Apr 2024

Immunostainer

Type:	Dako Omnis

Primary antibody

Clone:	L26
Producer:	Dako/Agilent
Product no. / lot no.:	M0755 / 41597660
Diluent:	Antibody Diluent
Diluent producer / no.:	Dako / K8006
Dilution factor:	1:200
Incubation time / temperature:	30 min. / 36°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Epredia Dewax and HIER Buffer H
Heating time at max. temp.:	20 min.
Maximum heating temp.:	97°C

Visualization system

Producer:	Dako Omnis
Product / no:	EnVision Flex / GV800/GV823
Linker:	Mouse LINKER
Incubation time linker:	10 min.
Incubation time polymer:	20 min.
Incubation temperature:	36°C

Chromogen

Producer:	Dako Omnis
Product / no:	DAB+ Substrate Chromogen System / GV825
Incubation time / temperature:	5 min. / 36°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.