

## Recommended protocol for PMS2

Obtained in run 71

06 Feb 2024

**Immunostainer**

Type: Dako Omnis

**Primary antibody**

Clone: EP51  
Producer: Dako/Agilent  
Product no. / lot no.: GA087 / 41569180  
Format: Ready-To-Use (prediluted)  
Incubation time / temperature: 20 min. / 24°C

**Epitope retrieval, HIER**

Device: On Board / On Machine  
Buffer: Dako Omnis Target Retrieval Solution, High pH  
Heating time at max. temp.: 30 min.  
Maximum heating temp.: 97°C

**Visualization system**

Producer: Dako Omnis  
Product / no: EnVision Flex / GV800/GV823  
Linker: Dual LINKER  
Incubation time linker: 10+10 min.  
Incubation time polymer: 20 min.  
Incubation temperature: 24°C

**Chromogen**

Producer: Dako Omnis  
Product / no: DAB+ Substrate Chromogen System / GV825  
Incubation time / temperature: 5 min. / 24°C  
Enhancement: CuSO4

**Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.