

Recommended protocol for CD20

Obtained in run 71

15 Feb 2024

Immunostainer

Type: MDStainer-Vitro

Primary antibody

Clone: L26
Producer: Master Diagnostica
Product no. / lot no.: MAD-002037QD/V / 20370053DS
Diluent: None
Diluent producer / no.: None / None
Dilution factor: 1:1
Incubation time / temperature: 25 min. / 25°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Master Diagnostica TR1 high pH solution
Heating time at max. temp.: 30 min.
Maximum heating temp.: 101°C

Visualization system

Producer: Master Diagnostica
Product / no: Master Polymer Plus Detection System (peroxidase) / MA
Linker: Primary Antibodies Amplifier Master
Incubation time linker: 13 min.
Incubation time polymer: 15 min.
Amplifier: Master Polymer Plus HRP
Incubation time amplifier: NA min.
Incubation temperature: 25°C

Chromogen

Producer: Master Diagnostica
Product / no: Master Polymer Plus Detection System (peroxidase) / MA
Incubation time / temperature: 10 min. / 25°C
Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided



Nordic Immunohistochemical Quality Control

Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.