Nordic Immunohistochemical Quality Control
Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CK-PAN

Obtained in run

13 Jul 2016

Immunostainer

Type:	Autostainer, LabVision

Primary antibody

Clone:	BS5
Producer:	Nordic Biosite
Product no. / lot no.:	BSH-7124-100 / BSH1-4t
Diluent:	Antibody Diluent
Dilution factor:	1:200
Incubation time / temperature:	30 min. / 20°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.:	20 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Nordic Biosite
Product / no:	NS / KDB-10007
Incubation time polymer:	30 min.
Incubation temperature:	20°C

Chromogen

Producer:	Nordic Biosite
Product / no:	DAB / BCB-20032
Incubation time / temperature:	10 min. / 20°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.