Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD20

Obtained in run 71

10 May 2024

Immunostainer

Type:	Tissue-Tek Genie Advanced Stai

Primary antibody

Clone:	L26
Producer:	Sakura Finetek
Product no. / lot no.:	8259-C010 / 8259B04023
Diluent:	Ab diluent
Diluent producer / no.:	Sakura Finetek / Ab diluent
Dilution factor:	1:0
Incubation time / temperature:	30 min min. / 20°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Sakura Finetek Tissue-Tek Genie High pH Antigen Retriev
Heating time at max. temp.:	45 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Sakura Finetek
Product / no:	Tissue-Tek Genie Pro Detection Kit, DAB / 8826-K250 / 88
Linker:	Tissue-Tek Genie Pro Detection Kit, DAB
Incubation time linker:	15 min.
Incubation time polymer:	20 min.
Incubation temperature:	20°C

Chromogen

Producer:	Tissue-Tek Genie PRO
Product / no:	Tissue-Tek Genie PRO Detection Kit, DAB / 8826-K250
Incubation time / temperature:	30 min. / 25°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.