Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for CD8

Obtained in run 72

10 Oct 2024

Immunostainer

Type:	Sakura Tissue-Tek Genie

Primary antibody

Clone:	C8/144B
Producer:	Sakura Finetek
Product no. / lot no.:	8252-C010 / 8252A15824
Format:	Ready-To-Use (prediluted)
Incubation time / temperature:	30 min. / 20°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Sakura Tissue-Tek Genie High pH Antigen Retrieval
Heating time at max. temp.:	45 min.
Maximum heating temp.:	98°C

Visualization system

Producer:	Sakura
Product / no:	Tissue-Tek Genie Pro Detection Kit, DAB / 8826-K250
Incubation time linker:	15 min.
Incubation time polymer:	20 min.
Incubation temperature:	20°C

Chromogen

Producer:	Sakura
Product / no:	Tissue-Tek Genie Pro Detection Kit, DAB / 8826-K250
Incubation time / temperature:	6 min. / 25°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.