**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for p16

Obtained in run 72

25 Jul 2024

**Immunostainer**

| | |
|---|---|
| Type: | Ventana Benchmark Ultra Plus |

**Primary antibody**

| | |
|---|---|
| Clone: | MX007 |
| Producer: | Medac/Zeta |
| Product no. / lot no.: | Z2016 / 230302673B |
| Diluent: | Antibody Diluent |
| Diluent producer / no.: | Ventana / 251-018 |
| Dilution factor: | 1:100 |
| Incubation time / temperature: | 24 min. / 36°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Ventana CC1 |
| Heating time at max. temp.: | 32 min. |
| Maximum heating temp.: | 90°C |

**Visualization system**

| | |
|---|---|
| Producer: | Ventana |
| Product / no: | OptiView DAB IHC Detection Kit / 760-700 |
| Incubation time linker: | 8 min. |
| Incubation time polymer: | 8 min. |
| Incubation temperature: | 36°C |

**Chromogen**

| | |
|---|---|
| Producer: | Ventana |
| Product / no: | OptiView DAB IHC Detection Kit / 760-700 |
| Incubation time / temperature: | 8 min. / 36°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.