Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for p16

Obtained in run 72

24 Sep 2024

Immunostainer

Type:	Dako Omnis

Primary antibody

Clone:	JC8
Producer:	Agilent
Product no. / lot no.:	GA783 / 41642448
Diluent:	None
Diluent producer / no.:	(None) / None
Dilution factor:	1:0
Incubation time / temperature:	20 min. / 32°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Dako Omnis Target Retrieval Solution, High pH
Heating time at max. temp.:	30 min.
Maximum heating temp.:	99°C

Visualization system

Producer:	Dako Omnis
Product / no:	EnVision Flex / GV800/GV823/GV900
Linker:	Mouse LINKER
Incubation time linker:	10 min.
Incubation time polymer:	20 min.
Incubation temperature:	32°C

Chromogen

Producer:	Dako Omnis
Product / no:	DAB+ Substrate Chromogen System / GV825
Incubation time / temperature:	5 min. / 32°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.