**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for CK19

Obtained in run 50

23 Mar 2017

**Immunostainer**

| | |
|---|---|
| Type: | Dako Autostainer Link 48 + |

**Primary antibody**

| | |
|---|---|
| Clone: | B170 |
| Producer: | Leica/novocastra |
| Product no. / lot no.: | NCL-CK19 / 6017257 |
| Diluent: | Antibody Diluent |
| Diluent producer / no.: | Dako / K8006 |
| Dilution factor: | 1:200 |
| Incubation time / temperature: | 20 min. / 23°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | PT-link / PT-module |
| Buffer: | Dako TRS High pH (3-1) |
| Heating time at max. temp.: | 10 min. |
| Maximum heating temp.: | 97°C |

**Visualization system**

| | |
|---|---|
| Producer: | Dako |
| Product / no: | EnVision FLEX+ / K8002/SM802 |
| Linker: | None |
| Incubation time polymer: | 20 min. |
| Incubation temperature: | 23°C |

**Chromogen**

| | |
|---|---|
| Producer: | Dako |
| Product / no: | DAB+ / DM827 / DM823 / K800 |
| Incubation time / temperature: | 10 min. / 23°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.