Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for HER2 IHC

Obtained in run B40

25 Sep 2025

Immunostainer

Type:	Dako Autostainer Link 48 +

Primary antibody

Clone:	HercepTest (polyclonal)
Producer:	Dako
Product no. / lot no.:	SK001 / 41846238
Format:	Ready-To-Use (prediluted)
Incubation time / temperature:	30 min. / 21°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Dako Herceptest
Heating time at max. temp.:	40 min.
Maximum heating temp.:	99°C

Visualization system

Producer:	Dako
Product / no:	Herceptest / SK001
Incubation time polymer:	30 min.
Incubation temperature:	21°C

Chromogen

Producer:	Dako
Product / no:	Herceptest / SK001
Incubation time / temperature:	10 min. / 21°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.