Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for PD-L1 (lung)

Obtained in run C3

01 Dec 2017

Immunostainer

Type:	Dako Autostainer Link 48 +

Primary antibody

Clone:	E1L3N
Producer:	Cell Signaling Technology
Product no. / lot no.:	13684 / 7
Diluent:	Antibody Diluent
Diluent producer / no.:	Dako / S3022
Dilution factor:	1:150
Incubation time / temperature:	40 min. / 22°C

Epitope retrieval, HIER

Device:	PT-link / PT-module
Buffer:	Dako TRS High pH (3-1)
Heating time at max. temp.:	20 min.
Maximum heating temp.:	95°C

Visualization system

Producer:	Dako
Product / no:	EnVision FLEX / K8000/SM802
Linker:	Linker, Rabbit
Incubation time linker:	15 min.
Incubation time polymer:	20 min.
Incubation temperature:	22°C

Chromogen

Producer:	Dako
Product / no:	DAB+ / K3467 / K3468
Incubation time / temperature:	10 min. / 22°C
Enhancement:	None

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.